

D Brash & Sons Ltd

37 Stamperland Crescent Clarkston, Glasgow, G76 8LH
 sales@brash-scales.com www.brash-scales.com



No: 26128

Certificate of Calibration

Nolan Recycling Ltd
 Waste Transfer Station
 The Old Quarry
 Stormy Down
 Bridgend
 CF33 4RS

Customer Ref: Weighbridge
 Location: Site
 Manufacturer: Leon
 Type: LD5218
 Serial No: L07340153
 Capacity: 50000 Kg x 20 Kg
 Next Due: April 2018

All test weights are traceable to the National Physical Standard as certified by a UKAS Accredited Laboratory

Test Weight Details

Set Number	Certificate Number
PB21-41	TR10038

Deviation / Tolerance for Compliance +/-

0 Kg - 10000 Kg	+/- 20 Kg
10000 Kg - 40000 Kg	+/- 40 Kg
40000 Kg - 50000 Kg	+/- 60 Kg

AS Found Hysteresis

Load	Loading	Deviation
0 Kg	0	0 Kg
6000 Kg	6000	0 Kg
12000 Kg	12000	0 Kg
20000 Kg	20000	0 Kg
34960 Kg	34960	0 Kg
44960 Kg	44960	0 Kg

Post Adjustment Hysteresis

Load	Loading	Deviation
0 Kg	0	0 Kg
6000 Kg	6000	0 Kg
12000 Kg	12000	0 Kg
20000 Kg	20000	0 Kg
34960 Kg	34960	0 Kg
44960 Kg	44960	0 Kg


As Found Repeatability

Zero	Load 1	Zero	Load 2	Zero	Load 3
0	44960	0	44960	0	44960

Post Adjustment Repeatability

Zero	Load 1	Zero	Load 2	Zero	Load 3
0	44960	0	44960	0	44960

Engineers Comments



As Found Eccentric Loading

Reading	Dev	Reading	Dev
1 6300	0 Kg	2 6300	0 Kg
3 6300	0 Kg	4 6300	0 Kg
5		6	
7		8	

Post Adjustment Eccentric Loading

Reading	Dev	Reading	Dev
1 6300	0 Kg	2 6300	0 Kg
3 6300	0 Kg	4 6300	0 Kg
5		6	
7		8	

The above detailed equipment has been tested in accordance with the UKWF Guidelines and found to be within specified tolerances unless otherwise stated

Issuing Branch: Cardiff

Date of Issue: Wed, 19 Apr 2017