

# CERTIFICATE OF CALIBRATION

DATE OF ISSUE: 2025-05-08 09:48:43

CERTIFICATE NUMBER: DBSN050039

ISSUED BY D BRASH & SONS LTD



**D Brash & Sons Ltd**  
37 Stamperland Crescent  
Clarkston, Glasgow, G76 8LH  
Tel: 0141 638 2284  
sales@dbrash.co.uk



<b>Customer</b> Nolan Recycling Ltd, Waste Transfer Station, The Old Quarry, , Stormy Down, Bridgend, CF33 4RS <b>Contact</b> Emma Cull	<b>Calibration Site</b> Nolan Recycling Ltd, Waste Transfer Station, The Old Quarry, , Stormy Down, Bridgend, CF33 4RS	<b>Approved Signatory</b>  <i>Ian Harvey</i>  Ian Harvey
<b>Model</b> Leon LD5218 <b>Serial No</b> L07340153	<b>Customer Ref</b> <b>Location</b>	<b>Capacity</b> 50 000kg <b>Division</b> 20kg
<b>Calibration Type</b> As Found <b>Calibrator</b> Ian Harvey	<b>Date of Calibration</b> Thu 08/May/2025 <b>Next Calibration Due</b> May 2026	<b>Internal Calibration Weight</b> N/A <b>Activated</b>
<b>Test Weight Sets</b> CA-Test unit		
<b>Comments</b> scale in tolerance.		

<b>As Found Eccentricity Test</b> Applied Load: 7 000kg					<b>Post Adjustment Eccentricity Test</b> Applied Load: 7 000kg			
Ref	Reading (kg)	Ref	Reading (kg)		Ref	Reading (kg)	Ref	Reading (kg)
1	7 000	4	7 000		1	7 000	4	7 000
2	6 980	5	7 000		2	6 980	5	7 000
3	7 000	6	7 000		3	7 000	6	7 000

<b>As Found Linearity</b>		
Applied Load (kg)	Reading (kg)	Difference (kg)
0	0	0
10 000	10 000	0
16 000	16 000	0
20 000	20 000	0
27 380	27 380	0
41 440	41 440	0

<b>Post Adjustment Linearity</b>		
Applied Load (kg)	Reading (kg)	Difference (kg)
0	0	0
10 000	10 000	0
16 000	16 000	0
20 000	20 000	0
27 380	27 380	0
41 440	41 440	0

<b>As Found Repeatability Test</b>		
Applied Load: 7 000kg		Zero Offset Load: 0kg
Unloaded (kg)	Loaded (kg)	Difference (kg)
0	7 000	7 000
0	7 000	7 000
0	7 000	7 000

<b>Post Adjustment Repeatability Test</b>		
Applied Load: 7 000kg		Zero Offset Load: 0kg
Unloaded (kg)	Loaded (kg)	Difference (kg)
0	7 000	7 000
0	7 000	7 000
0	7 000	7 000

This certificate provides traceability of measurement to the SI system of units and/or units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of D Brash & Sons Ltd.