

# CERTIFICATE OF CALIBRATION

DATE OF ISSUE: 2026-05-27 16:59:44


CERTIFICATE NUMBER: DBSN083188

ISSUED BY D BRASH & SONS LTD



**D Brash & Sons Ltd**  
37 Stamperland Crescent  
Clarkston, Glasgow, G76 8LH  
Tel: 0141 638 2284  
sales@dbrash.co.uk




<b>Customer</b> Nolan Recycling Ltd, Waste Transfer Station, The Old Quarry, , Stormy Down, Bridgend, CF33 4RS <b>Contact</b> Emma Cull	<b>Calibration Site</b> Nolan Recycling Ltd, Waste Transfer Station, The Old Quarry, , Stormy Down, Bridgend, CF33 4RS	<b>Approved Signatory</b>  Ian Harvey
<b>Model</b> Leon LD5218 <b>Serial No</b> L07340153	<b>Customer Ref</b> <b>Location</b>	<b>Capacity</b> 50 000kg <b>Division</b> 20kg
<b>Calibration Type</b> As Found <b>Calibrator</b> Ian Harvey	<b>Date of Calibration</b> Wed 27/May/2026 <b>Next Calibration Due</b> May 2027	<b>Internal Calibration Weight</b> N/A <b>Activated</b>

**Test Weight Sets** CA-Test unit

## Comments

scaleneeded adjustment to bring it back in tolerance.

As Found Eccentricity Test					Post Adjustment Eccentricity Test			
Applied Load: 7 000kg					Applied Load: 7 000kg			
Ref	Reading (kg)	Ref	Reading (kg)	Ref	Reading (kg)	Ref	Reading (kg)	
1	7 000	4	7 000	1	7 000	4	7 000	
2	6 980	5	7 020	2	6 980	5	7 020	
3	7 020	6	7 000	3	7 020	6	7 000	

As Found Linearity		
Applied Load (kg)	Reading (kg)	Difference (kg)
0	0	0
10 000	10 000	0
16 000	16 000	0
20 000	20 080	80
31 380	31 460	80
41 420	41 500	80

Post Adjustment Linearity		
Applied Load (kg)	Reading (kg)	Difference (kg)
0	0	0
10 000	10 000	0
16 000	16 000	0
20 000	20 000	0
31 380	31 380	0
41 420	41 420	0

As Found Repeatability Test		
Applied Load: 7 000kg		Zero Offset Load: 0kg
Unloaded (kg)	Loaded (kg)	Difference (kg)
0	7 000	7 000
0	7 000	7 000
0	7 000	7 000

Post Adjustment Repeatability Test		
Applied Load: 7 000kg		Zero Offset Load: 0kg
Unloaded (kg)	Loaded (kg)	Difference (kg)
0	7 000	7 000
0	7 000	7 000
0	7 000	7 000

This certificate provides traceability of measurement to the SI system of units and/or units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of D Brash & Sons Ltd.