

CERTIFICATE OF CALIBRATION

DATE OF ISSUE: 2024-04-18 16:09:35

CERTIFICATE NUMBER: DBSN019470

ISSUED BY D BRASH & SONS LTD

D Brash & Sons Ltd

37 Stamperland Crescent
Clarkston, Glasgow, G76 8LH


Tel: 0141 638 2284

sales@dbrash.co.uk



Customer Nolan Recycling Ltd, Waste Transfer Station, The Old Quarry, , Stormy Down, Bridgend, CF33 4RS Contact Emma Cull	Calibration Site Nolan Recycling Ltd, Waste Transfer Station, The Old Quarry, , Stormy Down, Bridgend, CF33 4RS	Approved Signatory  Mike Blackborow
Model Leon LD5218 Serial No L07340153	Customer Ref Location	Capacity 50 000kg Division 20kg
Calibration Type As Found Calibrator Mike Blackborow	Date of Calibration Thu 18/Apr/2024 Next Calibration Due April 2025	Internal Calibration Weight N/A Activated
Test Weight Sets CA-Test unit		

Comments

As Found Eccentricity Test					Post Adjustment Eccentricity Test			
Applied Load: 7 000kg					Applied Load: 7 000kg			
Ref	Reading (kg)	Ref	Reading (kg)		Ref	Reading (kg)	Ref	Reading (kg)
1	7 000	4	6 980		1	7 000	4	6 980
2	7 000	5	7 000	2	7 000	5	7 000	
3	7 000	6	7 000	3	7 000	6	7 000	

As Found Linearity		
Applied Load (kg)	Reading (kg)	Difference (kg)
0	0	0
10 000	10 000	0
16 000	16 000	0
20 000	19 980	-20
31 340	31 320	-20
41 340	41 320	-20

Post Adjustment Linearity		
Applied Load (kg)	Reading (kg)	Difference (kg)
0	0	0
10 000	10 000	0
16 000	16 000	0
20 000	19 980	-20
31 340	31 320	-20
41 340	41 320	-20

As Found Repeatability Test		
Applied Load: 7 000kg		Zero Offset Load: 0kg
Unloaded (kg)	Loaded (kg)	Difference (kg)
0	7 000	7 000
0	7 000	7 000
0	7 000	7 000

Post Adjustment Repeatability Test		
Applied Load: 7 000kg		Zero Offset Load: 0kg
Unloaded (kg)	Loaded (kg)	Difference (kg)
0	7 000	7 000
0	7 000	7 000
0	7 000	7 000

This certificate provides traceability of measurement to the SI system of units and/or units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of D Brash & Sons Ltd.